



RPM Error Allowance

Champion F-33D Series Centrifuge

It is suggested, but not required, for Ample Scientific Champion series centrifuges to be inspected in **2-year intervals**. The accuracy of these centrifuges should remain at **±5% of maximum rpm**. For example, a Champion F-33D centrifuge that spins 3,300 rpm should have an accuracy of ±165 rpm. Customers are strongly encouraged to consult Ample Scientific if abnormal behavior is observed during centrifugation.

If you have any questions regarding your centrifuge, please contact Ample Scientific

via phone at (678) 380-0369 or

via e-mail at sales@amplescientific.com

Thank you for choosing Ample Scientific. We appreciate your business.

Ample Scientific Customer Service Team



Ample Scientific LLC
4220 Steve Reynolds Blvd. Suite 9
Norcross, GA 30093 USA
(678) 380-0369
www.AmpleScientific.com
Sales@amplescientific.com